

Moving Class I & II Scales

Effective Date: January 30, 2014

[Back to Table of Contents](#)

 [NIST Handbook 44 G-UR.4.1., Sec. 2.20 UR.2.1.](#)

Policy

When should Class I & II scales be recalibrated?

NIST Handbook 44 gives specific requirements for the installation, operation and maintenance of weighing devices. Given the precise function and design of Class I & II scales, it is a requirement of the Wisconsin Bureau of Weights & Measures, once a weighing device has experienced a physical change of address it shall be recalibrated by licensed service company technician. This requirement includes dealers traveling from location to location purchasing precious metals.