

APPENDIX N BULK MILK TANKER SCREENING TEST FORM

**CHARM® ROSA SULF TEST
(Raw Commingled Cow Milk)
IMS #9-C18**

[Unless otherwise stated all tolerances are ±5%]

GENERAL REQUIREMENTS

1. See Appendix N General Requirements (App. N GR) items 1-8 & 15 _____

SAMPLES

2. See App. N GR item 9 _____

APPARATUS & REAGENTS

3. **Equipment** _____

- a. Charm Sciences Strip Incubator:
56±1°C 8 min timer – same incubator as SL beta-lactam test;
56±1°C Charm EZ display when message “Add milk to strip and close door” _____

1. Clean, properly maintained and located on a level surface _____

2. Check temperature daily (day of use); maintain records _____

- a. Charm EZ printout acceptable for daily temperature check (annual accuracy check required); maintain records _____

3. Temperature measuring device for each incubator (App N. GR item 3) _____

4. Lid closed (slightly sprung so that timer not active)-when not running tests _____

5. Incubator Temperature: _____

6. Timer if not included in incubator
Incubation Time of internal timer: _____

- b. ROSA® Reader, ROSA Pearl Reader (with or without ROSA Barcode option), Charm EZ or Charm Sciences equivalent with print out or download of data; manual available _____

Serial Number: _____

1. ROSA Reader V1.03 or higher (or if ROSA Pearl Reader or Charm EZ see 3.b.2) _____

a. Calibrators - 2 line for ROSA SULF Performance on SL beta-lactam channel acceptable _____

Range(s)

Result

Low: _____

High: _____

b. Maintain records _____

2. ROSA Pearl Reader V3.00 or higher or Charm EZ _____

a. Calibrators - Low and High for use in all assay channels _____

Range(s)

Result

Low: _____

(darker magenta)

High: _____

(lighter pink)

b. Maintain records _____

3. Calibrator serial numbers match reader SN _____

4. **Do not proceed if out of range.** Manufacturer should be contacted for corrective actions _____

5. Printer or computer link for hardcopy download _____

c. Pipettors - 300 μ L and disposable tips (see App. N GR item 7) _____

1. **FOR SCREENING ONLY** - Single use 300 μ L ROSA-pipet with overflow bulb to accurately measure amount of sample, supplied by manufacturer _____

4. Test Strips _____

a. Test Strips (EZ Compatible for Charm EZ) _____

Lot #: _____

Exp. Date: _____

QC Date: _____

By: _____

5. Sample and control agitation

- a. Mix milk sample(s)/control(s) 25 times in 7 sec with a 1 ft movement or vortex for 10 sec at maximum setting; use within 3 min. (samples/controls must be in appropriate containers to allow the use of vortexing)

6. Reagent Stability and Preparation

- 3. Charm ROSA- SULF reagents must be received within 72 hours if shipped non-refrigerated; over 72 hours must be refrigerated.

- b. Store reagents at 0.0-4.5°C, desiccant blue, maintain no longer than manufacturers expiration date.

- c. Negative Control

- 1. Previously tested sulfonamide negative raw milk.
- 2. Milk can be screened (previously tested) by the testing location making and or using the controls
- 3. Negative control must test -600 or more negative

Sample ID: _____ Test Value: _____

Date tested: _____

- 4. Use within 72 hours when maintained at 0.0-4.5°C
- 5. Or, aliquot within 24 hours and freeze at -15°C or colder in a non-frost-free freezer or in an insulated foam container in a frost-free freezer; use within 2 months

Lab Prep. Date: _____ Lab Exp. Date: _____

- a. Thaw slowly overnight in refrigerator or more rapidly in cold water. Mix well until sample is homogeneous

- 1. **Do not use if there is visible protein precipitation**

- b. Store at 0.0-4.5°C and use within 24 hours; do not refreeze

- 6. Day of use must produce -600 or more negative; maintain records

Do not proceed if out of range

d. Positive Control _____

1. Manufacturer supplied, do not use after manufacturer's expiration date _____

2. Lyophilized or tablet 10 ppb Sulfamethazine _____

Lot # _____ Exp. Date: _____ _____

3. Reconstitute 1 tablet with 5 mL Negative Control (raw milk) item 6.c _____

4. Tested +400 or more positive, used within 48 hours when maintained at 0.0-4.5°C _____

Lab Prep. Date: _____ Lab Exp. Date: _____ _____

5. Or, aliquot within 24 hours and freeze at -15°C or colder in a non-frost-free freezer or in an insulated foam container in a frost-free freezer; use within 1 month _____

Lab Prep. Date: _____ Lab Exp. Date: _____ _____

a. Thaw slowly overnight in refrigerator or more rapidly in cold water. Mix well until sample is homogeneous _____

1. **Do not use if there is visible protein precipitation** _____

b. Store at 0.0-4.5°C and use within 24 hours; do not refreeze _____

6. Day of use, must produce +400 or greater reading; maintain records _____

Test Value: _____ _____

Do not proceed if out of range _____

TECHNIQUE

7. Daily Performance and Operation Check _____

a. See App. N GR items 10.b-d _____

b. If using ROSA reader Versions 1.05 and higher, or ROSA-Pearl, use ESC 5 reader function to enter performance monitoring mode of reader; if using Charm EZ, use Menu to enter Performance Monitoring mode and "Perf Mon" to enter daily performance check; refer to manual for directions _____

c. Check Calibrators; items 3.b.1 or 3.b.2 _____

d. Positive and negative controls must give appropriate readings prior to any sample analysis (see App. N GR item 10.a) _____

e. Controls in-range when in performance monitoring mode, ROSA reader version 1.05 and higher, ROSA Pearl or Charm EZ

f. **Do not proceed if out of range**

8. Test Procedure

a. Set out required number of test strips and place them in a dry labeled container at room temperature, or take out strips as needed

1. Discard unused test strips at the end of the day

b. Label test strips, one for each test sample and each control. Avoid crushing sample compartment(s)

c. Mix milk sample(s)/control(s) (See Item 5 a)

d. Place strip into incubator

e. While holding strip flat, peel back plastic (to 'peel to here' line) to expose sample pad compartment. Avoid lifting the wick and sponge under tape

1. For multiple samples, complete items 8.d-g for each sample/control, before starting test of next sample

2. Complete all samples within 2 min of placing first strip in incubator

f. Add 300 μ L of mixed sample/control to corresponding strip

1. Using pipettor (item 3.c) with new tip for each sample/control, draw up 300 μ L avoiding foam or bubbles

a. Remove tip from liquid

b. While holding the pipettor vertically, expel test portion slowly into either side well of appropriate strip

2. **FOR SCREENING ONLY** - Using new manufacturer-provided ROSA-pipet (item 3.c.1) for each sample/control

a. Squeeze top bulb while holding vertically with bulb and overflow reservoir side pointing down, draw up test portion avoiding foam and bubbles. Sample should completely fill pipet shaft and overflow into the bottom half of the overflow reservoir

b. Remove tip from liquid

c. While holding the ROSA-pipet vertically, expel test portion slowly into either side well of appropriate strip. Excess portion should remain in reservoir

g. Re-seal plastic firmly around sample pad compartment

h. ROSA Reader and Charm EZ (read only mode) _____

1. Close lid and latch ROSA incubator to start automatic timer in the incubator. If no automatic timer in incubator, set external timer for 8 min. Incubate 8 min not to exceed 9 min. _____

2. At end of incubation visually inspect C (Control) line. An absent C line, a partial C line or an indistinct C line indicates an invalid test; and the sample/control must be re-tested _____

3. Insert only valid test(s) in reader _____

a. ROSA reader set to appropriate channel _____

1. SULF slow blink for ROSA SULF test _____

2. Press ENTER, reading and interpretation appear in 5 sec, read strips within 5 min of completion of incubation. Strips may be held vertically, sample compartment down while waiting to be read _____

b. Charm EZ automatically sets channel when color coded strip inserted _____

1. Close door; reading and interpretation appear in 5 sec, read strips within 5 min of completion of incubation. Strips may be held vertically, sample compartment down while waiting to be read _____

i. Charm EZ (incubate and read mode) _____

1. Charm EZ automatically sets channel and incubator temperature when color coded strip inserted. Optionally enter sample ID _____

2. Peel strip (8.e) and add milk (8.f) _____

3. Close door to begin _____

4. Charm EZ automatically prompts for further testing when positive _____

9. Interpretation with Reader _____

a. If there is a negative or zero reading on the reader, sample is a Negative (NF) _____

b. If there is a positive reading on the reader, sample is an Initial Positive _____

10. Verification of Initial Positive Tanker Samples Done at the Same Testing Facility (see App. N GR item 11) _____

a. Set out four test strips and label as negative control, positive control, and two strips with the initial positive sample ID. Avoid crushing sample compartment(s) _____

- b. Mix milk sample(s)/control(s) (See Item 5 a) _____
- c. Place a labeled strip into incubator. _____
- d. While holding strip flat, peel back plastic (to 'peel to here' line) to expose sample pad compartment. Avoid lifting the wick and sponge under tape _____
 - 1. Complete items 10.c-f for each sample/control, before starting test of next sample/control _____
 - 2. Complete pipetting all samples within 2 min of placing first strip in incubator _____
- e. Add 300 μ L of mixed sample/control to corresponding strip _____
 - 1. Using pipettor (item 3.c) with new tip for each sample/control, draw up 300 μ L avoiding foam or bubbles _____
 - a. Remove tip from liquid _____
 - b. While holding the pipettor vertically, expel test portion slowly into either side well of appropriate strip _____
- f. Re-seal plastic firmly around sample pad compartment _____
- g. ROSA Reader and Charm EZ (must confirm in read only mode) _____
 - 1. Close lid and latch ROSA incubator to start automatic timer in the incubator. If no automatic timer in incubator, set external timer for 8 min. Incubate 8 min not to exceed 9 min _____
 - 2. At end of incubation remove strips from incubator. Strips may be held vertically with sample compartment down. Visually inspect each strip C (Control) line. An absent C line, a partial C line or an indistinct C line indicates an invalid test; and the sample/control must be re-tested _____
 - 3. Insert only valid test(s) in reader _____
 - a. ROSA reader set to appropriate channel _____
 - 1. SULF slow blink for ROSA SULF test _____
 - 2. Press ENTER, reading and interpretation appear in 5 sec, read strips within 5 min of completion of incubation. Strips may be held vertically, sample compartment down while waiting to be read _____
 - 3. Positive and negative controls must give appropriate readings prior to any sample analysis (see items 6.c-d) _____
 - a. Controls in-range when in performance monitoring mode, ROSA reader version 1.05 and higher, ROSA Pearl _____

- b. Charm EZ automatically sets channel when color coded strip inserted _____
 - 1. From initial positive screen, press Continue or Confirm Later _____
 - a. If press Confirm Later, then call the confirmation routine back by pressing Main Menu. _____
 - b. Select Perf Mon. then from next menu select CNFM-Later button _____
 - c. Select initial positive sample ID and click ok _____
 - 2. Follow Prompts to insert negative and then positive control. Click Continue Confirmation upon successful completion of positive control. _____
 - 3. Follow prompts to enter the first duplicate sample. Insert the strip and then close door; reading and interpretation appear in 5 sec. Follow prompts to enter the second reading and interpretation appear in 5 sec. Read strips within 5 min of completion of incubation. Strips may be held vertically, sample compartment down while waiting to read _____
 - 4. Charm EZ automatically determines Presumptive Positive or Not Found. _____
- 11. Confirmation of Presumptive Positive Tanker Samples (see App. N GR item 12) [Only in an accredited laboratory or by a CIS] _____**
- 12. Traceback of Producer(s) on a Confirmed Positive Tanker (see App. N GR item 13) [Only in an accredited laboratory or by a CIS (refer to M-a-85 current revision for a listing of test kits to assure equivalence)] _____**
- 13. Re-instatement of Producer(s) [Only in an accredited laboratory or by a CIS (refer to M-a-85 current revision for a listing of test kits to assure equivalence)] _____**
- 14. Reporting (see App. N GR item 14) _____**